

Status: Currently Official on 14-Feb-2025  
Official Date: Official Prior to 2013  
Document Type: Reagents  
DocId: GUID-A4BE08AD-55B4-40BF-A9D8-6E075CC86A6D\_1\_en-US  
DOI: [https://doi.org/10.31003/USPNF\\_R1788\\_01\\_01](https://doi.org/10.31003/USPNF_R1788_01_01)  
DOI Ref: 23529

© 2025 USPC  
Do not distribute

# 1-Chloroadamantane,

$C_{10}H_{15}Cl$  170.68 CAS RN<sup>®</sup>: 935-56-8.—White crystalline solid.

**Assay:** Inject an appropriate volume into a gas chromatograph (see [Chromatography \(621\)](#)) equipped with a flame-ionization detector, helium being used as the carrier gas. The following conditions have been found suitable: a 0.25-mm × 30-m capillary column coated with a 1- $\mu$ m layer of phase G2; the injection port temperature is maintained at 250°; the detector temperature is maintained at 300°; and the column temperature is maintained at 150° and programmed to rise 10° per minute to 280°. The area of the  $C_{10}H_{15}Cl$  peak is not less than 97.5% of the total peak area.

**Auxiliary Information** - Please [check for your question in the FAQs](#) before contacting USP.

Topic/Question	Contact	Expert Committee
1-CHLOROADAMANTANE	<a href="#">Margareth R.C. Marques</a> Principal Scientific Liaison	HDQ Headquarters

**Most Recently Appeared In:**

Pharmacopeial Forum: Volume No. Information currently unavailable

**Current DocID:** [GUID-A4BE08AD-55B4-40BF-A9D8-6E075CC86A6D\\_1\\_en-US](#)

**DOI:** [https://doi.org/10.31003/USPNF\\_R1788\\_01\\_01](https://doi.org/10.31003/USPNF_R1788_01_01)

**DOI ref:** [23529](#)

OFFICIAL