Status: Currently Official on 14-Feb-2025
Official Date: Official as of 01-May-2020
Document Type: NF Monographs
DocId: GUID-A1442242-9127-41FF-8C52-321A00ABBFF7_2_en-US
DOI: https://doi.org/10.31003/USPNF_M11120_02_01
DOI Ref: vw91g

© 2025 USPC Do not distribute

Butylparaben

Portions of the monograph text that are national *USP* text, and are not part of the harmonized text, are marked with symbols (*) to specify this fact.

 $C_{11}H_{14}O_3$

194.23

Benzoic acid, 4-hydroxy-, butyl ester;

Butyl p-hydroxybenzoate CAS RN®: 94-26-8.

DEFINITION

Butylparaben contains NLT 98.0% and NMT 102.0% of $C_{11}H_{14}O_3$.

IDENTIFICATION

Change to read:

- A. Spectroscopic Identification Tests (197), Infrared Spectroscopy: 197M (CN 1-May-2020)
- B. Melting Range or Temperature (741): 68°-71°

ASSAY

• Procedure

Mobile phase, Sample solution, Standard solution B, and **Chromatographic system:** Proceed as directed in the procedure for *Related Substances*.

System suitability

Sample: Standard solution B **Suitability requirements**

Relative standard deviation: NMT 0.85% for six injections

Analysis

Samples: Sample solution and Standard solution B

Calculate the percentage of Butylparaben in the Sample solution:

Result =
$$P \times (r_{II} \times C_S)/(r_S \times C_{II})$$

P = labeled purity of <u>USP Butylparaben RS</u> expressed as a percentage

 r_{ij} = peak area of butylparaben from the Sample solution

 C_S = concentration of butylparaben in Standard solution B (mg/mL)

r_s = peak area of butylparaben from *Standard solution B*

 C_{ij} = concentration of Butylparaben in the Sample solution (mg/mL)

Acceptance criteria: 98.0%-102.0%

IMPURITIES

- RESIDUE ON IGNITION (281): NMT 0.1%, determined on a 1.0-g sample
- RELATED SUBSTANCES

Mobile phase: Methanol and a 6.8 g/L solution of potassium dihydrogen phosphate (1:1 v/v)

Sample solution: Dissolve 50.0 mg of Butylparaben in 2.5 mL of methanol, and dilute with *Mobile phase* to 50.0 mL. Dilute 10.0 mL of this solution with *Mobile phase* to 100.0 mL.

Standard solution A: 5.0 µg/mL each of p-hydroxybenzoic acid, <u>USP Propylparaben RS</u>, and <u>USP Butylparaben RS</u> in Mobile phase

Standard solution B: Dissolve 50.0 mg of <u>USP Butylparaben RS</u> in 2.5 mL of methanol, and dilute with *Mobile phase* to 50.0 mL. Dilute 10.0 mL of this solution with *Mobile phase* to 100.0 mL.

Standard solution C: Dilute 1.0 mL of the Sample solution with Mobile phase to 20.0 mL. Dilute 1.0 mL of this solution with Mobile phase to

Standard solution D: 50 μ g/mL of iso-butylparaben in *Mobile phase* **Standard solution E:** *Standard solution D* in *Standard solution B* (1 in 100)

Chromatographic system

(See Chromatography (621), System Suitability.)

Mode: LC

Detector: UV 272 nm

Column: 4.6-mm × 15-cm; 5-µm packing L1

Column temperature: 35° Flow rate: 1.3 mL/min Injection volume: 10 µL

Run time: About 1.5 times the retention time of butylparaben

System suitability

Sample: Standard solutions A and E

[Note—The retention time of butylparaben is about 22 min; the relative retention times for *p*-hydroxybenzoic acid, propylparaben, and isobutylparaben with a reference to butylparaben are about 0.1, 0.5, and 0.9 min, respectively.]

Suitability requirements

Resolution: NLT 5.0 between the propylparaben and butylparaben peaks from *Standard solution A* and NLT 1.5 between the isobutylparaben and butylparaben peaks from *Standard solution E*

Analysis

Samples: Sample solution and Standard solution C

[Note—Disregard any limit that is 0.2 times the area of the principal peak from Standard solution C (0.1%).]

Acceptance criteria

p-Hydroxybenzoic acid: The peak area from the *Sample solution*, multiplied by 1.4 to correct for the calculation of content, is NMT the area of the principal peak from *Standard solution C* (0.5%).

Unspecified impurities: The peak area of each impurity from the *Sample solution* is NMT the area of the principal peak from *Standard solution C* (0.5%).

Total impurities: The total peak area for all impurities from the *Sample solution* is NMT twice the area of the principal peak from *Standard solution C* (1.0%).

SPECIFIC TESTS

• Acidity: To 2 mL of Butylparaben solution prepared in the Color of Solution test add 3 mL of alcohol, 5 mL of carbon dioxide-free water, and 0.1 mL of bromocresol green TS. Titrate with 0.10 N sodium hydroxide.

Acceptance criteria: NMT 0.1 mL is required to produce a blue color.

Color of Solution

Butylparaben solution: Dissolve 1 g in alcohol, and dilute with alcohol to 10 mL.

Acceptance criteria: This solution is clear and not more intensely colored than alcohol or a solution prepared immediately before use by mixing 2.4 mL of ferric chloride CS, 1.0 mL of cobaltous chloride CS, and 0.4 mL of cupric sulfate CS with 0.3 N hydrochloric acid to make 10 mL, and diluting 5 mL of this solution with 0.3 N hydrochloric acid to make 100 mL. Make the comparison by viewing the solutions downward in matched color-comparison tubes against a white surface (see *Color and Achromicity* (631)).

ADDITIONAL REQUIREMENTS

- PACKAGING AND STORAGE: Preserve in well-closed containers.
- USP REFERENCE STANDARDS (11)

USP Butylparaben RS
USP Propylparaben RS

 $\textbf{Auxiliary Information} \cdot \textbf{Please} \ \underline{\textbf{check for your question in the FAQs}} \ \textbf{before contacting USP}.$

Topic/Question	Contact	Expert Committee
BUTYLPARABEN	Documentary Standards Support	SE2020 Simple Excipients

Chromatographic Database Information: Chromatographic Database

Most Recently Appeared In:

Pharmacopeial Forum: Volume No. PF 34(6)

Current DocID: GUID-A1442242-9127-41FF-8C52-321A00ABBFF7_2_en-US

DOI: https://doi.org/10.31003/USPNF_M11120_02_01

DOI ref: vw91g

