

Status: Currently Official on 12-Feb-2025
 Official Date: Official as of 01-May-2021
 Document Type: General Chapter
 DocId: GUID-6DE27279-F886-4535-8B6A-DCAF584C6C12_3_en-US
 DOI: https://doi.org/10.31003/USPNF_M933_03_01
 DOI Ref: 7er9k

© 2025 USPC
 Do not distribute

(644) CONDUCTIVITY OF SOLUTIONS

Portions of the chapter text that are national *USP* text and are not part of the harmonized text, are marked with symbols (♦) to specify this fact.

INTRODUCTION

This chapter provides information on how to apply electrical conductivity measurements (hereafter referred to as “conductivity”) of fluid solutions, including pure fluids. ♦This chapter does not replace the official [Water Conductivity \(645\)](#) procedure, which is used to ensure the ionic purity of compendial waters such as [Water for Injection](#), [Purified Water](#), [Pure Steam](#) condensate, and [Sterile Water for Injection](#), among others. ♦ This chapter is intended for ♦other ♦ fluid applications when conductivity is used to measure, monitor, or control chemical dispensing, chemical purity, ionic concentration, and other applications where the ionic character of the fluid needs to be known or controlled.

Applications include, but are not limited to, solutions that may be used in clean-in-place, chromatography detection, ionic solution preparations, endpoint detection, dosing, fermentation, and buffer production. In some cases, conductivity measurements can be extended to pure organic fluids such as alcohols and glycols where a weak conductivity signal exists, and the signal can be significantly increased if the organics become contaminated with water or salts. ♦[NOTE—For additional background information, see [Theory and Practice of Electrical Conductivity Measurements of Solutions \(1644\)](#).]

Conductivity is the measurement of the ability of a fluid to conduct electricity via its chemical ions. The ability of any ion to electrically conduct is directly related to its ion mobility. Conductivity is directly proportional to the concentrations of ions in the fluid, according to *Equation 1*:

$$\kappa = 1000 \sum_i^{all\ ions} C_i \lambda_i$$

κ = conductivity (S/cm)

C_i = concentration of chemical ion *i* (mol/L)

λ_i = specific molar conductance of ion *i* (S · cm²/mol)

Although the SI unit S/m is the appropriate SI unit for conductivity, historically the unit S/cm has been selected by industry as the accepted unit.

On the basis of *Equation 1*, conductivity is not ion selective because it responds to all ions. Furthermore, the specific molar conductance of each ion is different. As a result, unless the percentage composition of ions of the solution is limited and known, the precise concentrations of ionic species cannot be determined from conductivity measurements. However, for examples such as a solution of a single salt or acid or base, such as a caustic solution used in cleaning, the precise concentration can be directly determined. Despite the lack of ionic specificity, conductivity is a valuable laboratory and process tool for measurement and control of total ionic content because it is proportional to the sum of the concentrations of all ionic species (anions and cations) for diluted solutions as described in *Equation 1*. At higher concentrations, conductivity measurements are not perfectly linear with concentration. Conductivity measurements cannot be applied to solids or gases, but they can be applied to the condensate of gases.

Another variable that influences conductivity measurements is the fluid temperature. As the fluid temperature increases, the ion conductance increases, making this physicochemical phenomenon the predominant reason for the temperature-compensation requirement when testing conductive fluids.

The conductivity, κ , is proportional to the conductance, *G* (S), of a fluid between two electrodes (*Equation 2*):

$$\kappa = G \times \left(\frac{d}{A} \right) = G \times K$$

κ = conductivity (S/cm)

G = conductance (S)

d = distance between the electrodes (cm)

A = area of the conducting electrodes (cm²)

K = cell constant (cm^{-1}), which also equals the ratio of d/A

The resistivity ρ ($\Omega\text{-cm}$) of the fluid is, by definition, the reciprocal of the conductivity (Equation 3):

$$\rho = 1/\kappa = 1/(G \times K) = R/K$$

ρ = resistivity ($\Omega\text{-cm}$)

κ = conductivity (S/cm)

G = conductance (S)

K = cell constant (cm^{-1})

R = resistance (Ω), which is the reciprocal of the conductance, G

APPARATUS

An electrical conductivity measurement consists of the determination of resistance of the fluid between and around the electrodes of the conductivity sensor. To achieve this measurement, the primary instrumentation is the resistance-measuring circuit and the conductivity sensor, and they are usually connected by a cable when the sensor and the user interface are separated.

The resistance measurement is made by applying an alternating current (AC, meaning the flow of electric charge periodically reverses direction) voltage (or current) to the electrodes, measuring the current (or voltage), and calculating the resistance according to Ohm's law. The alternating source is used to prevent the polarization (collection of ions) at the electrodes. Depending on the instrument, the measuring frequency of the measuring system adjusts automatically according to the measuring conditions of the instrument, and there may be multiple resistance-measuring circuits embedded in the measuring system. The resistance-measurement circuit may be embedded in the transmitter or in the sensor.

The conductivity sensor consists of at least two electrical conductors of a fixed size and geometry, separated by an electrical insulator. The electrodes, insulator, and any other wetted materials should be constructed of materials that are unreactive to fluids with which they may come into contact. Also, the sensor construction should withstand the environmental conditions (process or ambient temperature, pressure, cleaning applications) that it would be subjected to.

Most conductivity sensors have temperature devices such as a platinum resistance temperature device (RTD) or negative temperature coefficient (NTC) thermistor embedded inside the sensor, although external temperature measurement is possible. The purpose of the temperature measurement is for temperature compensation of the conductivity measurement.

CELL CONSTANT DETERMINATION

The purpose of the sensor's cell constant is to normalize the conductance (or resistance) measurement for the geometrical construction of the two electrodes.

The cell constant is determined by immersing the conductivity sensor in a solution of known conductivity. Solutions of known conductivity can be obtained by preparation of specific mixtures according to national authoritative sources or procurement of commercially available certified and traceable standard solutions. These recipes or certified solutions can range from 5 to 200,000 $\mu\text{S/cm}$, depending on the level of accuracy desired. Alternatively the cell constant is determined by comparison to other reference conductivity measuring systems (also available as an accredited calibration service). [NOTE—Conductivity measurements are not perfectly linear with concentration.]

The measured cell constant of the conductivity sensor must be within 5% of the nominal value indicated by the sensor certificate, unless otherwise prescribed.

Modern conductivity sensors normally do not change their cell constant over their lifetime. If a change of the cell constant is detected during calibration, a cleaning of the sensor is appropriate according to the manufacturer's recommendations. Following that, the calibration procedure should be repeated. Sometimes "memory effects" appear, particularly when changing from high to low concentrations if the sensor is not well flushed.

CALIBRATION OF TEMPERATURE

In addition to verifying the sensor's cell constant, the embedded temperature device (or external temperature device) should be appropriately calibrated for the application to apply the temperature compensation algorithm accurately. The temperature accuracy that is required depends on the criticality of the temperature to the application. An accuracy of $\pm 1^\circ$ typically suffices.

CALIBRATION OF MEASUREMENT ELECTRONICS

The measurement circuit of the system is fundamentally an AC resistance measuring device. Appropriate verification and/or calibration of the measuring circuit is required for measurement systems with signal transfer via analog cable. This is accomplished by disconnecting the measuring circuit from the sensor's electrodes, attaching traceable resistors of known value with the cable of the measurement system to the measuring circuit, and verifying that the measured resistance agrees with the resistor value to an acceptable level. A typical acceptance criterion for the resistance accuracy is $<2\%$ of the reading at resistances $>100 \Omega$, and increasing to 5% at lower resistances. However, the application criticality should ultimately determine the desired accuracy.

For conductivity systems that cannot have the resistance-measuring circuit disconnected from the electrodes (e.g., measurement circuit and electrodes in one mutual housing), it may be difficult to directly adjust or verify the circuit accuracy, depending on the sensor design. An alternative method of verifying the measurement system integrity is a system calibration according to the procedures for the cell constant determination for each measuring circuit that is intended to be used.

If verification/calibration of the sensor's cell constant, temperature device, and measuring circuit are done at the same service interval, the measuring circuit should be verified first, the temperature device next, and the cell constant last. Because all of these parameters are typically very stable due to modern electronics and stable sensor construction, frequent calibration (such as daily) is not usually required. Comparison to qualified reference systems is also a proper means of calibration. Calibration is performed at appropriate intervals as defined in the quality management system.

Change to read:

TEMPERATURE COMPENSATION

Because the conductivity of a fluid is temperature dependent, temperature compensation of the conductivity measurement is necessary unless otherwise prescribed (e.g., purified water, water for injection). An appropriate temperature compensation algorithm will ensure that changes in the conductivity measurement can be ascribed to concentration changes and not temperature changes. Conductivity measurements are normally referenced to 25°. A common form of linear temperature compensation uses *Equation 4*:

$$\kappa_{25} = \frac{\kappa_T}{[1 + \alpha(T - 25)]}$$

κ_{25} = conductivity compensated to 25°

κ_T = conductivity at *T*

α = temperature coefficient of the conductivity

T = measured temperature

A temperature coefficient of 2.1% per 1° is commonly used for many salt solutions. Most salt-based solutions have linear compensation factors ranging from 1.9% to 2.2% per 1°. Depending on the fluid samples, other forms of temperature compensation may be appropriate. Non-linear temperature compensation data for a variety of solutions is widely available (e.g., as described in ISO 7888 *Water quality—Determination of electrical conductivity*). In cases of very low conductivity (<10 µS/cm), such as purified water for cleaning/rinsing purposes, (USP 1-May-2021) two compensations need to be made. One is for the intrinsic conductivity of water, and the other is for the other ionic species in water. These compensations are normally combined and embedded in the microprocessor-controlled conductivity measurement systems. This is not supplied in all conductivity measurement technologies.

Change to read:

CONDUCTIVITY MEASUREMENT OF FLUIDS

For off-line or at-line batch measurements, rinse the cleaned sensor with the fluid to be measured. Then immerse the sensor in the fluid to be measured, and record the temperature and the temperature-compensated conductivity as required. Be sure that the position of the sensor in the container does not affect the conductivity measurement, because the container walls can affect the measurement for some electrode designs.

For continuous on-line or in-line measurements, install the cleaned sensor into the pipe, tank, or other containment vessel, and flush, if necessary. Make sure proper installation procedures are applied to prevent bubbles or particles from collecting between the electrodes. Be sure that the position of the sensor in the pipe or tank does not affect the conductivity measurement, because the nearby surfaces can affect the measurement for some electrode designs.

Record the temperature and the temperature-compensated conductivity as required.

For all batch or continuous measurements, ensure that the wetted components of the sensor are compatible with the fluid and the temperature to be measured.

Auxiliary Information - Please [check for your question in the FAQs](#) before contacting USP.

Topic/Question	Contact	Expert Committee
<644> CONDUCTIVITY OF SOLUTIONS	Antonio Hernandez-Cardoso Senior Scientific Liaison	GCCA2020 General Chapters - Chemical Analysis 2020

Most Recently Appeared In:

Pharmacopeial Forum: Volume No. PF 41(5)

Current DocID: [GUID-6DE27279-F886-4535-8B6A-DCAF584C6C12_3_en-US](#)

DOI: https://doi.org/10.31003/USPNF_M933_03_01

DOI ref: [7er9k](#)