

Status: Currently Official on 16-Feb-2025
Official Date: Official Prior to 2013
Document Type: Reagents
DocId: GUID-32156200-F132-416C-9089-1A81438CCA45_1_en-US
DOI: https://doi.org/10.31003/USPNF_R1453_01_01
DOI Ref: 1rs8x

© 2025 USPC
Do not distribute

Nitric Oxide–Nitrogen Dioxide Detector Tube:

A fuse-sealed glass tube so designed that gas may be passed through it and containing suitable absorbing filters and support media for an oxidizing layer and the indicator diphenyl benzidine.

- **MEASURING RANGE:** 0.5 to 10 ppm.

[NOTE—Available from Draeger Safety, Inc., www.draeger.com, or from Gastec Corp., www.gastec.co.jp, distributed in the USA by www.nextteq.com.]

Auxiliary Information - Please [check for your question in the FAQs](#) before contacting USP.

Topic/Question	Contact	Expert Committee
NITRIC OXIDE–NITROGEN DIOXIDE DETECTOR TUBE	Margareth R.C. Marques Principal Scientific Liaison	HDQ Headquarters

Most Recently Appeared In:

Pharmacopeial Forum: Volume No. Information currently unavailable

Current DocID: [GUID-32156200-F132-416C-9089-1A81438CCA45_1_en-US](#)

DOI: https://doi.org/10.31003/USPNF_R1453_01_01

DOI ref: [1rs8x](#)

