

Status: Currently Official on 16-Feb-2025
Official Date: Official Prior to 2013
Document Type: NF Monographs
DocId: GUID-3FC6F6BE-7E7B-4511-8566-58F6C93E2703_1_en-US
DOI: https://doi.org/10.31003/USPNF_M56925_01_01
DOI Ref: us5oi

© 2025 USPC
Do not distribute

Nitrogen 97 Percent

DEFINITION

Nitrogen 97 Percent is Nitrogen produced from air by physical separation methods. It contains NLT 97.0%, by volume, of nitrogen (N₂).

IDENTIFICATION

- **A.** The flame of a burning wood splinter is extinguished when inserted into a test tube filled with Nitrogen 97 Percent. [NOTE—Exercise caution.]

ASSAY

- **PROCEDURE**

Standard: Oxygen–helium certified standard (see [Reagents, Indicators, and Solutions](#))

Sample: Nitrogen 97 Percent

Chromatographic system

(See [Chromatography \(621\), System Suitability](#).)

Mode: GC

Detector: Thermal conductivity

Column: 3-m length × 4-mm inside diameter: molecular sieve prepared from a synthetic alkali-metal aluminosilicate capable of absorbing molecules having diameters of up to 0.5 nm and completely separating oxygen from nitrogen

Carrier gas: Helium (99.99%)

Temperature: Thermostatically controlled

Analysis

Samples: Standard and Sample

Introduce the Samples separately into the gas chromatograph by means of a gas sampling valve.

Acceptance criteria: The peak response produced by the Sample exhibits a retention time corresponding to that produced by the Standard and is equivalent to NMT 3.0% of oxygen when compared to the peak response of the Standard, indicating NLT 97.0%, by volume, of nitrogen (N₂).

IMPURITIES

[NOTE—Reduce the container pressure by means of a regulator. Measure the gases with a gas volume meter downstream from the detector tube to minimize contamination or change of the specimens.]

- **CARBON DIOXIDE**

Sample: 1000 ± 50 mL

Analysis: Pass the Sample through a carbon dioxide detector tube (see [Reagents, Indicators, and Solutions](#)) at the rate specified for the tube.

Acceptance criteria: The indicator change corresponds to NMT 300 ppm

- **CARBON MONOXIDE**

Sample: 1000 ± 50 mL

Analysis: Pass the Sample through a carbon monoxide detector tube (see [Reagents, Indicators, and Solutions](#)) at the rate specified for the tube.

Acceptance criteria: NMT 10 ppm

- **SULFUR DIOXIDE**

Sample: 1000 ± 50 mL

Analysis: Pass the Sample through a sulfur dioxide detector tube (see [Reagents, Indicators, and Solutions](#)) at the rate specified for the tube.

Acceptance criteria: NMT 5 ppm

- **LIMIT OF NITRIC OXIDE AND NITROGEN DIOXIDE**

Sample: 500 ± 50 mL

Analysis: Pass the *Sample* through a nitric oxide–nitrogen dioxide detector tube (see [Reagents, Indicators, and Solutions](#)) at the rate specified for the tube.

Acceptance criteria: NMT 2.5 ppm

• **LIMIT OF OXYGEN**

Analysis: Determined as directed in the Assay

Acceptance criteria: NMT 3.0%

SPECIFIC TESTS

• **ODOR**

Analysis: Carefully open the container valve to produce a moderate flow of gas. Do not direct the gas stream toward the face, but deflect a portion of the stream toward the nose.

Acceptance criteria: No appreciable odor is discernible.

ADDITIONAL REQUIREMENTS

• **PACKAGING AND STORAGE:** Preserve in cylinders or in a low-pressure collecting tank.

• **LABELING:** Where it is piped directly from the collecting tank to the point of use, label each outlet "Nitrogen 97 Percent".

Auxiliary Information - Please [check for your question in the FAQs](#) before contacting USP.

Topic/Question	Contact	Expert Committee
NITROGEN 97 PERCENT	Documentary Standards Support	SM42020 Small Molecules 4
REFERENCE STANDARD SUPPORT	RS Technical Services RSTECH@usp.org	SM42020 Small Molecules 4

Chromatographic Database Information: [Chromatographic Database](#)

Most Recently Appeared In:

Pharmacopeial Forum: Volume No. PF 35(4)

Current DocID: [GUID-3FC6F6BE-7E7B-4511-8566-58F6C93E2703_1_en-US](#)

DOI: https://doi.org/10.31003/USPNF_M56925_01_01

DOI ref: [us5oi](#)