

Status: Currently Official on 16-Feb-2025
Official Date: Official Prior to 2013
Document Type: Reagents
DocId: GUID-FC2D74A6-8FCD-4BCD-A519-D39834256190_1_en-US
DOI: https://doi.org/10.31003/USPNF_R1458_01_01
DOI Ref: 1nhz1

© 2025 USPC
Do not distribute

Olefin Detector Tube:

A fuse-sealed glass tube so designed that gas may be passed through it and containing suitable absorbing filters and support media for the indicator in a stabilized form of permanganate.

- **MEASURING RANGE:** 0.06 to 3.2 Vol.-% Propylene, 0.04 to 2.4 Vol.-% Butylene.

[NOTE—Available from Draeger Safety, Inc., www.draeger.com.]

Auxiliary Information - Please [check for your question in the FAQs](#) before contacting USP.

| Topic/Question | Contact | Expert Committee |
|----------------------|--|------------------|
| OLEFIN DETECTOR TUBE | Margareth R.C. Marques Principal Scientific Liaison | HDQ Headquarters |

Most Recently Appeared In:

Pharmacopeial Forum: Volume No. Information currently unavailable

Current DocID: [GUID-FC2D74A6-8FCD-4BCD-A519-D39834256190_1_en-US](#)

DOI: https://doi.org/10.31003/USPNF_R1458_01_01

DOI ref: [1nhz1](#)

