

Status: Currently Official on 17-Feb-2025
Official Date: Official Prior to 2013
Document Type: Reagents
DocId: GUID-FC6C9EF4-FAC1-4D7D-8CB0-F09813112751_1_en-US
DOI: https://doi.org/10.31003/USPNF_R2547_01_01
DOI Ref: u6ug5

© 2025 USPC
Do not distribute

Water Vapor Detector Tube:

A fuse-sealed glass tube so designed that gas may be passed through it and containing suitable absorbing filters and support media for the indicator, which consists of a selenium sol in suspension in sulfuric acid.

- **MEASURING RANGE:** 5 to 250 mg per cubic meter.

[NOTE—Available from Draeger Safety, Inc., www.draeger.com, or from Gastec Corp., www.gastec.co.jp, distributed in the USA by www.nextteq.com.]

Auxiliary Information - Please [check for your question in the FAQs](#) before contacting USP.

Topic/Question	Contact	Expert Committee
WATER VAPOR DETECTOR TUBE	Margareth R.C. Marques Principal Scientific Liaison	HDQ Headquarters

Most Recently Appeared In:

Pharmacopeial Forum: Volume No. Information currently unavailable

Current DocId: GUID-FC6C9EF4-FAC1-4D7D-8CB0-F09813112751_1_en-US

DOI: https://doi.org/10.31003/USPNF_R2547_01_01

DOI ref: [u6ug5](#)

